

Application note 001

Cell counts and viability

Product description	The NucleoCounter™ system is comprised of the NucleoCounter™, NucleoCassettes™, Lysis buffer (Reagent A) and Stabilizing buffer (Reagent B). An optional part of the system is the NucleoView™ software.
Application	<p>The NucleoCounter system enables the user to obtain absolute cell counts (total and non-viable) and thereby determine the viability of cell suspensions from a wide range of cultured mammalian cells.</p> <p>The NucleoCounter is developed as a stand-alone instrument. Optionally the NucleoCounter can be connected to a computer using the NucleoView software, which offers a variety of features such as documentation of the results and calculation of viability.</p>
Principle	<p>In order to determine the total concentration of cells in a suspension of mammalian cells, a sample of the suspension is first treated with Lysis buffer and then Stabilizing buffer. The objective of the Lysis buffer is to permeate the plasma membranes, thereby allowing the nuclei to be stained with propidium iodide, which is coated on the inside of the NucleoCassette. The Lysis buffer also contributes to the disaggregation of cell clusters. The following treatment with Stabilizing buffer raises the pH value of the mixture in order to allow propidium iodide to stain the nuclei more efficiently. Approximately 50 µl of the stabilized lysate is then drawn into the NucleoCassette. The NucleoCassette is placed in the NucleoCounter where the cells are counted.</p> <p>To obtain the concentration of non-viable cells the cell suspension is loaded directly into the NucleoCassette without any pre-treatment with Lysis buffer or Stabilizing buffer. Thereby only the non-viable cells (with impaired plasma membranes) are stained with propidium iodide and counted.</p> <p>Using the NucleoCounter system it is possible to determine the total concentration of cells as well as the concentration of non-viable cells. Subsequently the viability may be calculated.</p>
Procedure	First the procedure for determining the total concentration of cells is stated and then the procedure for determining the concentration of non-viable cells. Finally the method for calculating the viability of the cells on basis of the two mentioned measurements is shown.

Total cell count

1. Mix 1 part of mammalian cells in suspension with 1 part of Lysis buffer (e.g. 200 µl cell suspension and 200 µl Lysis buffer).
2. Mix thoroughly (use a vortex mixer).
3. Add 1 part of Stabilizing buffer to the mixture, which is now called the stabilized lysate.
4. Mix thoroughly (use a vortex mixer).

Technology that counts

5. Draw a sample of the stabilized lysate by inserting the tip of the NucleoCassette into the lysate and pressing the piston.
6. Insert the NucleoCassette into the NucleoCounter, close the lid and press RUN. After approximately 30 seconds the total concentration of nuclei in the stabilized lysate is displayed on the NucleoCounter and on the computer if one is connected to the NucleoCounter. Note that it is the number of cells pr. ml in the stabilized lysate drawn into the NucleoCassette, which is displayed. Therefore, in order to calculate the number of mammalian cells pr. ml in the original suspension, the displayed number must be multiplied by the multiplication factor. The multiplication factor is 3 assuming the suspension of cells has not been diluted prior to treatment with Lysis buffer and Stabilizing buffer and the procedure above has been followed.

Non-viable cell count

1. Draw a sample of the suspension of mammalian cells into the NucleoCassette. The cells must not be treated with Lysis buffer and Stabilizing buffer prior to counting the number of non-viable cells.
2. Insert the NucleoCassette into the NucleoCounter, close the lid and press RUN. After approximately 30 seconds the concentration of non-viable cells in the suspension of cells is displayed on the NucleoCounter and the computer if one is connected to the NucleoCounter.

Note

To assure reliable cell counts with the NucleoCounter, it is recommended that the total concentration of cells in the stabilized lysate should be within the limits of $5 \cdot 10^3$ cells/ml to $2 \cdot 10^6$ cells/ml.

If the concentration of cells in the stabilized lysate is below $5 \cdot 10^3$ cells/ml then the concentration of cells per volume may be increased by centrifugation followed by resuspension of the pellet using PBS or growth media.

If the concentration of cells in the stabilized lysate is above $2 \cdot 10^6$ cells/ml, the suspension of cells can be diluted with PBS or growth media to achieve the desired concentration. Hereafter the suspension is treated with Lysis buffer and Stabilizing buffer.

Note

When counting the number of non-viable cells, the cells are not pretreated with Lysis buffer and Stabilizing buffer and they are thereby not diluted. Thus, the number displayed on the NucleoCounter and/or the computer is the actual number of non-viable cells pr. ml in the suspension. If the suspension of cells has been up-concentrated or diluted the appropriate multiplication factor must be used to calculate the correct concentration of non-viable cells in the original cell suspension.

Technology that counts

Viability

It is possible to calculate the viability of the cells after having obtained the total concentration of cells and the concentration of non-viable cells. Assuming that the cells have been counted as described above the calculation of viability is as follows.

$$\% \text{viability} = \frac{C_t \cdot M_t - C_{nv} \cdot M_{nv}}{C_t \cdot M_t} \cdot 100\%$$

$\% \text{viability}$ The percentage of viable cells in the original cell suspension.

C_t The total concentration of cells in the NucleoCassette (the displayed result of the total cell count).

C_{nv} The concentration of non-viable cells in the NucleoCassette (the result displayed when counting the non-viable cells).

M_t The multiplication factor used for the total cell count (most often 3).

M_{nv} The multiplication factor used for the non-viable cell count (most often 1).

Handling and storage

For handling and storage of reagents and NucleoCassettes refer to the individual packing labels.

Warnings and precautions

For safe handling and disposal of the reagents and NucleoCassettes refer to the packing labels and the NucleoCounter user's guide.

Limitations

The NucleoCounter system is for research use only. The results presented by the NucleoCounter system depend on correct use of the reagents, NucleoCassettes and the NucleoCounter. Refer to the NucleoCounter user's guide for instructions and limitations.

Disclaimer

ChemoMetec A/S reserves the right to introduce changes in the product to incorporate new technology. This application note is subject to change without notice.

Trademarks

NucleoCounter™, NucleoCassette™ and NucleoView™ are trademarks of ChemoMetec A/S.

Copyright

Copyright © ChemoMetec A/S 2002. All rights reserved. No part of this publication may be reproduced, stored in a retrieval system, or transmitted in any form or by any means, electronic, mechanical, photocopying, recording or otherwise, without the prior written consent of ChemoMetec A/S, Gydevang 43, DK-3450 Allerød, Denmark.

Technology that counts